

2020年03月ET測定結果

採取日		測定部位	2020/3/2	2020/3/3	2020/3/9	2020/3/11	2020/3/13	2020/3/23	2020/3/25	2020/3/27
測定場所		曜日	月	火	月	水	金	月	水	金
EU/mL		担当(本館)	豊田			稲富	吉岩	石井	中山	豊田
		担当(東館)	本田	本田	武田					
1	3-01	GC-110N				<0.001				
2	3-02	GC-110N				<0.001				
3	3-03	GC-110N				<0.001				
4	3-04	GC-110N				<0.001				
5	3-05	GC-300N●					<0.001			
6	3-06	GC-300N●					<0.001			
7	3-07	GC-300N●					<0.001			
8	3-08	GC-300N●					<0.001			
9	3-09	GC-300N●							<0.001	
10	3-10	GC-300N●							<0.001	
11	3-11	GC-300N●							<0.001	
12	3-12	GC-300N●							<0.001	
13	3-13	DBB-100NX								<0.001
14	3-14	DBB-100NX								<0.001
15	予備	DCS-27								<0.001
16	セントラル① ETRF入口	DAB-NX ①						<0.001		
17	セントラル① ETRF出口	DAB-NX ①						<0.001		
18	セントラル② ETRF入口	DAB-E ②						<0.001		
19	セントラル② ETRF出口	DAB-E ②						<0.001		
20	RO① EUF(奥側)	MIZ ①	<0.001							
21	RO① EUF(手前側)	MIZ ①	<0.001							
22	RO② EUF(奥側)	MIZ ②	<0.001							
23	RO② EUF(手前側)	MIZ ②	<0.001							
24	東2-03号機	DCS-100NX		<0.001						
25	東2-04号機	DCS-100NX		<0.001						
26	東2-05号機	DCS-100NX		<0.001						
27	東2-06号機	DCS-100NX		<0.001						
28	東2-07号機	DCS-100NX			<0.001					
29	東2-08号機	DCS-100NX			<0.001					
30	東2-09号機	DCS-100NX			<0.001					
31	東2-10号機	DCS-100NX			<0.001					
32	東RO モジュール	東MIZ	<0.001							
33	東RO タンク後	東MIZ	<0.001							
34	東RO EUF①	東MIZ	<0.001							
35	東RO EUF②	東MIZ	<0.001							
	合計	35	8	4	4	4	4	4	4	3

2020年03月生菌測定結果

採取日		装置	2020/3/2	2020/3/3	2020/3/9	2020/3/11	2020/3/13	2020/3/23	2020/3/25	2020/3/27
計測日		2020/4/15	2020/3/9	2020/3/10	2020/3/16	2020/3/18	2020/3/20	2020/3/30	2020/4/1	2020/4/3
培養温度		25～30℃	月	火	月	水	金	月	水	金
測定場所		担当（本館）	豊田			稲富	吉岩	石井	中山	豊田
		担当（東棟）	本田	本田	武田					
1	3-01	GC-110N				0/50ml				
2	3-02	GC-110N				0/50ml				
3	3-03	GC-110N				0/50ml				
4	3-04	GC-110N				0/50ml				
5	3-05	GC-300N●					0/50ml			
6	3-06	GC-300N●					0/50ml			
7	3-07	GC-300N●					0/50ml			
8	3-08	GC-300N●					0/50ml			
9	3-09	GC-300N●							0/50ml	
10	3-10	GC-300N●							0/50ml	
11	3-11	GC-300N●							0/50ml	
12	3-12	GC-300N●							0/50ml	
13	3-13	DBB-100NX								0/50ml
14	3-14	DBB-100NX								0/50ml
15	予備	DCS-27								0/50ml
16	セントラル① ETRF入口	DAB-NX ①						0/50ml		
17	セントラル① ETRF出口	DAB-NX ①						0/50ml		
18	セントラル② ETRF入口	DAB-E ②						0/50ml		
19	セントラル② ETRF出口	DAB-E ②						0/50ml		
20	RO① EUF(奥側)	MIZ ①	0/50ml							
21	RO① EUF(手前側)	MIZ ①	0/50ml							
22	RO② EUF(奥側)	MIZ ②	0/50ml							
23	RO② EUF(手前側)	MIZ ②	0/50ml							
24	東2-03号機	DCS-100NX		0/50ml						
25	東2-04号機	DCS-100NX		0/50ml						
26	東2-05号機	DCS-100NX		0/50ml						
27	東2-06号機	DCS-100NX		0/50ml						
28	東2-07号機	DCS-100NX			0/50ml					
29	東2-08号機	DCS-100NX			0/50ml					
30	東2-09号機	DCS-100NX			0/50ml					
31	東2-10号機	DCS-100NX			0/50ml					
32	東RO モジュール	東MIZ	0/50ml							
33	東RO タンク後	東MIZ	0/50ml							
34	東RO EUF①	東MIZ	0/50ml							
35	東RO EUF②	東MIZ	0/50ml							
合計		35	8	4	4	4	4	4	4	3