

2020年04月ET測定結果

採取日		測定部位	2020/4/1	2020/4/2	2020/4/6	2020/4/8	2020/4/10	2020/4/20	2020/4/22	2020/4/24
測定場所		曜日	水	木	月	水	金	月	水	金
EU/mL		担当(本館)			尾ノ上	森川	本田	本田	石井	中山
		担当(東館)	武田	武田						
1	7-01	DCS-100NX				<0.001				
2	7-02	DCS-100NX				<0.001				
3	7-03	DCS-100NX				<0.001				
4	7-04	DCS-100NX				<0.001				
5	7-05	DCS-100NX					<0.001			
6	7-06	DCS-100NX					<0.001			
7	7-07	DCS-100NX					<0.001			
8	7-08	DCS-100NX					<0.001			
9	7-09	DCS-100NX							<0.001	
10	7-10	DCS-100NX							<0.001	
11	7-11	DCS-100NX							<0.001	
12	7-12	DCS-100NX							<0.001	
13	7-13	DCS-100NX								<0.001
14	7-14	DCS-100NX								<0.001
15	7-15	DCS-100NX								<0.001
16	7-16	DCS-100NX								<0.001
17	セントラル③ ETRF入口	DAB-NX ③			<0.001					
18	セントラル③ ETRF出口	DAB-NX ③			<0.001					
19	セントラル④ ETRF入口	DAB-NX ④			<0.001					
20	セントラル④ ETRF出口	DAB-NX ④			<0.001					
21	RO① モジュール	MIZ ①						<0.001		
22	RO② モジュール	MIZ ①						<0.001		
23	RO① タンク後	MIZ ②						<0.001		
24	RO② タンク後	MIZ ②						<0.001		
25	東2-11号機	DCS-100NX	<0.001							
26	東2-12号機	DCS-100NX	<0.001							
27	東3-01号機	GC-X01	<0.001							
28	東3-02号機	GC-X01	<0.001							
29	東4-08号機	DCS-100NX		<0.001						
30	東セントラル① ETRF入口	東DAB-E ①		<0.001						
31	東セントラル① ETRF出口1	東DAB-E ①		<0.001						
32	東セントラル① ETRF出口2	東DAB-E ①		<0.001						
合計		32	4	4	4	4	4	4	4	4

2020年04月生菌測定結果

採取日		装置	2020/4/1	2020/4/2	2020/4/6	2020/4/8	2020/4/10	2020/4/20	2020/4/22	2020/4/24
計測日		2020/5/4	2020/4/8	2020/4/9	2020/4/13	2020/4/15	2020/4/17	2020/4/27	2020/4/29	2020/5/1
培養温度		25～30℃	水	木	月	水	金	月	水	金
測定場所		担当（本館）			尾ノ上	森川	本田	本田	石井	中山
		担当（東棟）	武田	武田						
1	7-01	DCS-100NX				0/50ml				
2	7-02	DCS-100NX				0/50ml				
3	7-03	DCS-100NX				0/50ml				
4	7-04	DCS-100NX				0/50ml				
5	7-05	DCS-100NX					0/50ml			
6	7-06	DCS-100NX					0/50ml			
7	7-07	DCS-100NX					0/50ml			
8	7-08	DCS-100NX					0/50ml			
9	7-09	DCS-100NX							0/50ml	
10	7-10	DCS-100NX							0/50ml	
11	7-11	DCS-100NX							0/50ml	
12	7-12	DCS-100NX							0/50ml	
13	7-13	DCS-100NX								0/50ml
14	7-14	DCS-100NX								0/50ml
15	7-15	DCS-100NX								0/50ml
16	7-16	DCS-100NX								0/50ml
17	セントラル③ ETRF入口	DAB-NX ③			0/50ml					
18	セントラル③ ETRF出口	DAB-NX ③			0/50ml					
19	セントラル④ ETRF入口	DAB-NX ④			0/50ml					
20	セントラル④ ETRF出口	DAB-NX ④			0/50ml					
21	RO① モジュール	MIZ ①						0/50ml		
22	RO② モジュール	MIZ ①						6/50ml		
23	RO① タンク後	MIZ ②						0/50ml		
24	RO② タンク後	MIZ ②						0/50ml		
25	東2-11号機	DCS-100NX	0/50ml							
26	東2-12号機	DCS-100NX	0/50ml							
27	東3-01号機	GC-X01	0/50ml							
28	東3-02号機	GC-X01	0/50ml							
29	東4-08号機	DCS-100NX		0/50ml						
30	東セントラル① ETRF入口	東DAB-E ①		0/50ml						
31	東セントラル① ETRF出口1	東DAB-E ①		0/50ml						
32	東セントラル① ETRF出口2	東DAB-E ①		0/50ml						
合計		32	4	4	4	4	4	4	4	4