

2020年06月ET測定結果

採取日		測定部位	2020/6/1	2020/6/2	2020/6/8	2020/6/17	2020/6/19	2020/6/22	2020/6/24	2020/6/26
測定場所		曜日	月	火	月	水	金	月	水	金
EU/mL		担当(本館)			中山	渡辺	山田	光山	石井	吉岩
		担当(東館)	都	都						
1	6-01	DBB-100NX							<0.001	
2	6-02	DBB-100NX							<0.001	
3	6-03	DBB-100NX							<0.001	
4	6-04	DBB-100NX							<0.001	
5	6-05	DCS-100NX								<0.001
6	6-06	DCS-100NX								<0.001
7	6-07	DCS-100NX								<0.001
8	6-08	DCS-100NX								<0.001
9	6-09	DCS-100NX				<0.001				
10	6-10	DCS-100NX				<0.001				
11	6-11	DCS-100NX				<0.001				
12	6-12	DCS-100NX				<0.001				
13	6-13	DCS-100NX					<0.001			
14	6-14	DCS-100NX					<0.001			
15	6-15	DCS-100NX					<0.001			
16	6-16	DCS-100NX					<0.001			
17	セントラル③ ETRF入口	DAB-NX ③			<0.001					
18	セントラル③ ETRF出口	DAB-NX ③			<0.001					
19	セントラル④ ETRF入口	DAB-NX ④			<0.001					
20	セントラル④ ETRF出口	DAB-NX ④			<0.001					
21	RO① モジュール	MIZ ①						<0.001		
22	RO② モジュール	MIZ ①						<0.001		
23	RO① タンク後	MIZ ②						<0.001		
24	RO② タンク後	MIZ ②						<0.001		
25	東3-11号機	DCS-100NX		<0.001						
26	東3-12号機	DBB-27		<0.001						
27	東4-10号機	DCS-100NX		<0.001						
28	東予備機	DCS-100NX		<0.001						
29	東RO モジュール	東MIZ	<0.001							
30	東RO タンク後	東MIZ	<0.001							
31	東RO EUF①	東MIZ	<0.001							
32	東RO EUF②	東MIZ	<0.001							
	合計	32	4	4	4	4	4	4	4	4

2020年06月生菌測定結果

採取日		装置	2020/6/1	2020/6/2	2020/6/8	2020/6/17	2020/6/19	2020/6/22	2020/6/24	2020/6/26
計測日		2020/7/3	2020/6/8	2020/6/9	2020/6/15	2020/6/24	2020/6/26	2020/6/29	2020/7/1	2020/7/3
培養温度		25～30℃	月	火	月	水	金	月	水	金
測定場所		担当（本館）			中山	渡辺	山田	光山	石井	吉岩
		担当（東棟）	都	都	都					
1	6-01	DBB-100NX							0/50ml	
2	6-02	DBB-100NX							0/50ml	
3	6-03	DBB-100NX							0/50ml	
4	6-04	DBB-100NX							0/50ml	
5	6-05	DCS-100NX								0/50ml
6	6-06	DCS-100NX								0/50ml
7	6-07	DCS-100NX								0/50ml
8	6-08	DCS-100NX								0/50ml
9	6-09	DCS-100NX				0/50ml				
10	6-10	DCS-100NX				0/50ml				
11	6-11	DCS-100NX				0/50ml				
12	6-12	DCS-100NX				0/50ml				
13	6-13	DCS-100NX					0/50ml			
14	6-14	DCS-100NX					0/50ml			
15	6-15	DCS-100NX					0/50ml			
16	6-16	DCS-100NX					0/50ml			
17	セントラル③ ETRF入口	DAB-NX ③			0/50ml					
18	セントラル③ ETRF出口	DAB-NX ③			0/50ml					
19	セントラル④ ETRF入口	DAB-NX ④			0/50ml					
20	セントラル④ ETRF出口	DAB-NX ④			0/50ml					
21	RO① モジュール	MIZ ①						0/50ml		
22	RO② モジュール	MIZ ①						13/50ml		
23	RO① タンク後	MIZ ②						0/50ml		
24	RO② タンク後	MIZ ②						0/50ml		
25	東3-11号機	DCS-100NX		0/50ml						
26	東3-12号機	DBB-27		0/50ml						
27	東4-10号機	DCS-100NX		0/50ml						
28	東予備機	DCS-100NX		0/50ml						
29	東RO モジュール	東MIZ	13/50ml		0/50ml					
30	東RO タンク後	東MIZ	0/50ml							
31	東RO EUF①	東MIZ	0/50ml							
32	東RO EUF②	東MIZ	0/50ml							
	合計	33	4	4	5	4	4	4	4	4