

2020年09月ET測定結果

採取日		測定部位	2020/9/1	2020/9/2	2020/9/3	2020/9/4	2020/9/14	2020/9/16	2020/9/18	2020/9/21	2020/9/30
測定場所		曜日	火	水	木	金	月	水	金	月	水
EU/mL		担当(本館)		山田		石井		吉岩	大塚	森川	渡辺
		担当(東館)	都	武田	武田		武田				
1	1-01	DCS-100NX						<0.001			
2	1-02	DCS-100NX						<0.001			
3	1-03	DCS-100NX						<0.001			
4	1-04	DCS-100NX						<0.001			
5	2-05	GC-X01							<0.001		
6	2-06	GC-X01							<0.001		
7	2-07	GC-X01							<0.001		
8	2-08	GC-X01							<0.001		
9	2-09	GC-X01									<0.001
10	2-10	GC-X01									<0.001
11	2-11	GC-X01									<0.001
12	2-12	GC-X01									<0.001
13	9-09	GC-X01		<0.001							
14	9-10	GC-X01		<0.001							
15	9-11	GC-X01		<0.001							
16	9-12	GC-X01		<0.001							
17	9-13	GC-X01				<0.001					
18	予備	DCS-100NX				<0.001					
19	セントラル① ETRF入口	DAB-NX ①								<0.001	
20	セントラル① ETRF出口	DAB-NX ①								<0.001	
21	セントラル② ETRF入口	DAB-NX ②								<0.001	
22	セントラル② ETRF出口	DAB-NX ②								<0.001	
23	東2-03号機	DCS-100NX	<0.001								
24	東2-04号機	DCS-100NX	<0.001								
25	東2-05号機	DCS-100NX	<0.001								
26	東2-06号機	DCS-100NX	<0.001								
27	東2-07号機	DCS-100NX		<0.001							
28	東2-08号機	DCS-100NX		<0.001							
29	東2-09号機	DCS-100NX		<0.001							
30	東2-10号機	DCS-100NX		<0.001							
31	東RO モジュール	東MIZ			<0.001		再検				
32	東RO タンク後	東MIZ			<0.001						
33	東RO EUF①	東MIZ			<0.001						
34	東RO EUF②	東MIZ			<0.001						
	合計	34	4	8	4	2	0	4	4	4	4

2020年09月生菌測定結果

採取日		装置	2020/9/1	2020/9/2	2020/9/3	2020/9/4	2020/9/14	2020/9/16	2020/9/18	2020/9/21	2020/9/30
計測日		2020/10/8	2020/9/8	2020/9/9	2020/9/10	2020/9/11	2020/9/21	2020/9/23	2020/9/25	2020/9/28	2020/10/7
培養温度		25～30℃	火	水	木	金	月	水	金	月	水
測定場所		担当（本館）		山田		石井		吉岩	大塚	森川	渡辺
		担当（東棟）	都	武田	武田		武田				
1	1-01	DCS-100NX						0/50ml			
2	1-02	DCS-100NX						0/50ml			
3	1-03	DCS-100NX						0/50ml			
4	1-04	DCS-100NX						0/50ml			
5	2-05	GC-X01							0/50ml		
6	2-06	GC-X01							0/50ml		
7	2-07	GC-X01							0/50ml		
8	2-08	GC-X01							0/50ml		
9	2-09	GC-X01									0/50ml
10	2-10	GC-X01									0/50ml
11	2-11	GC-X01									0/50ml
12	2-12	GC-X01									0/50ml
13	9-09	GC-X01		0/50ml							
14	9-10	GC-X01		0/50ml							
15	9-11	GC-X01		0/50ml							
16	9-12	GC-X01		0/50ml							
17	9-13	GC-X01				0/50ml					
18	予備	DCS-27				0/50ml					
19	セントラル① ETRF入口	DAB-NX ①								0/50ml	
20	セントラル① ETRF出口	DAB-NX ①								0/50ml	
21	セントラル② ETRF入口	DAB-E ②								0/50ml	
22	セントラル② ETRF出口	DAB-E ②								0/50ml	
23	東2-03号機	DCS-100NX	0/50ml								
24	東2-04号機	DCS-100NX	0/50ml								
25	東2-05号機	DCS-100NX	0/50ml								
26	東2-06号機	DCS-100NX	0/50ml								
27	東2-07号機	DCS-100NX		0/50ml							
28	東2-08号機	DCS-100NX		0/50ml							
29	東2-09号機	DCS-100NX		0/50ml							
30	東2-10号機	DCS-100NX		0/50ml							
31	東RO モジュール	東MIZ			42/50ml		0/50ml				
32	東RO タンク後	東MIZ			0/50ml						
33	東RO EUF①	東MIZ			0/50ml						
34	東RO EUF②	東MIZ			0/50ml						
	合計	35	4	8	4	2	1	4	4	4	4