

2020年12月生菌測定結果

| 採取日 | 装置 | 2019/12/30 | 2019/12/31 | 2020/12/1 | 2020/12/5 | 2020/12/7 | 2020/12/9 | 2020/12/10 | 2020/12/21 | 2020/12/23 | 2020/12/25 | 2020/12/28 |
|------|------------------|------------|------------|-----------|------------|------------|------------|------------|------------|------------|------------|------------|
| 計測日 | 2021/1/4 | 2020/1/6 | 2020/1/7 | 2020/12/8 | 2020/12/12 | 2020/12/14 | 2020/12/16 | 2020/12/17 | 2020/12/28 | 2020/12/30 | 2021/1/1 | 2021/1/4 |
| 培養温度 | 25~30℃ | 月 | 火 | 火 | 土 | 月 | 水 | 木 | 月 | 水 | 金 | 月 |
| 測定場所 | 担当(本館) | 豊田 | 渡辺 | | | 吉岩 | 光山 | 中山 | 山田 | 森川 | 中山 | 大塚 |
| | 担当(東棟) | | | 都 | 武田 | 都 | | | | | | |
| 1 | 7-01 | DCS-100NX | | | | | 0/50ml | | | | | |
| 2 | 7-02 | DCS-100NX | | | | | 0/50ml | | | | | |
| 3 | 7-03 | DCS-100NX | | | | | 0/50ml | | | | | |
| 4 | 7-04 | DCS-100NX | | | | | 0/50ml | | | | | |
| 5 | 7-05 | DCS-100NX | | | | | | 0/50ml | | | | |
| 6 | 7-06 | DCS-100NX | | | | | | 0/50ml | | | | |
| 7 | 7-07 | DCS-100NX | | | | | | 0/50ml | | | | |
| 8 | 7-08 | DCS-100NX | | | | | | 0/50ml | | | | |
| 9 | 7-09 | DCS-100NX | | | | | | | 0/50ml | | | |
| 10 | 7-10 | DCS-100NX | | | | | | | 0/50ml | | | |
| 11 | 7-11 | DCS-100NX | | | | | | | 0/50ml | | | |
| 12 | 7-12 | DCS-100NX | | | | | | | 0/50ml | | | |
| 13 | 7-13 | DCS-100NX | | | | | | | | 0/50ml | | |
| 14 | 7-14 | DCS-100NX | | | | | | | | 0/50ml | | |
| 15 | 7-15 | DCS-100NX | | | | | | | | 0/50ml | | |
| 16 | 7-16 | DCS-100NX | | | | | | | | 0/50ml | | |
| 17 | 9-09 | GC-X01 | | 0/50ml | | | | | | | | |
| 18 | 9-10 | GC-X01 | | 0/50ml | | | | | | | | |
| 19 | 9-11 | GC-X01 | | 0/50ml | | | | | | | | |
| 20 | 9-12 | GC-X01 | | 0/50ml | | | | | | | | |
| 21 | セントラル③ ETRF入口 | DAB-NX ③ | | | | 0/50ml | | | | | | |
| 22 | セントラル③ ETRF出口 | DAB-NX ③ | | | | 0/50ml | | | | | | |
| 23 | セントラル④ ETRF入口 | DAB-NX ④ | | | | 0/50ml | | | | | | |
| 24 | セントラル④ ETRF出口 | DAB-NX ④ | | | | 0/50ml | | | | | | |
| 25 | RO① モジュール | MIZ ① | | | | | | | 10/50ml | | | |
| 26 | RO② モジュール | MIZ ① | | | | | | | 15/50ml | | | |
| 27 | RO① タンク後 | MIZ ② | | | | | | | 0/50ml | | | |
| 28 | RO② タンク後 | MIZ ② | | | | | | | 0/50ml | | | |
| 29 | RO① EUF(奥側) | MIZ ① | 0/50ml | | | | | | | | | 0/50ml |
| 30 | RO① EUF(手前側) | MIZ ① | 0/50ml | | | | | | | | | 0/50ml |
| 31 | RO② EUF(奥側) | MIZ ② | 0/50ml | | | | | | | | | 0/50ml |
| 32 | RO② EUF(手前側) | MIZ ② | 0/50ml | | | | | | | | | 0/50ml |
| 33 | 東3-11号機 | DCS-100NX | | | 0/50ml | | | | | | | |
| 34 | 東3-12号機 | DBB-27 | | | 0/50ml | | | | | | | |
| 35 | 東4-10号機 | DCS-100NX | | | 0/50ml | | | | | | | |
| 36 | 東予備機 | DCS-100NX | | | 0/50ml | | | | | | | |
| 37 | 東RO モジュール | 東MIZ | | | 28/50ml | 0/50ml | | | | | | |
| 38 | 東RO タンク後 | 東MIZ | | | 0/50ml | | | | | | | |
| 39 | 東RO EUF① | 東MIZ | | | 0/50ml | | | | | | | |
| 40 | 東RO EUF② | 東MIZ | | | 0/50ml | | | | | | | |
| | 合計 | 45 | 4 | 4 | 4 | 4 | 5 | 4 | 4 | 4 | 4 | 4 |