

2021年05月生菌測定結果

採取日	装置	2021/5/1	2021/5/10	2021/5/12	2021/5/13	2021/5/14	2021/5/19	2021/5/24	2021/5/26	2021/5/28
計測日	2021/6/11	2021/5/8	2021/5/17	2021/5/19	2021/5/20	2021/5/21	2021/5/26	2021/5/31	2021/6/2	2021/6/4
培養温度	25~30℃	土	月	水	木	金	水	月	水	金
測定場所	担当(本館)		森川			光山	渡辺	中山	中山	豊田
	担当(東棟)	都		武田	武田					
1	1-01	DCS-100NX					0/50ml			
2	1-02	DCS-100NX					0/50ml			
3	1-03	DCS-100NX					0/50ml			
4	1-04	DCS-100NX					0/50ml			
5	1-05	DCS-100NX				0/50ml				
6	1-06	DCS-100NX				0/50ml				
7	1-07	DCS-100NX				0/50ml				
8	1-08	DCS-100NX				0/50ml				
9	1-09	DCS-100NX							0/50ml	
10	1-10	DCS-100NX							0/50ml	
11	1-11	DCS-100NX							0/50ml	
12	1-12	DCS-100NX							0/50ml	
13	1-13	DCS-100NX								0/50ml
14	1-14	DCS-100NX								0/50ml
15	1-15	DCS-100NX								0/50ml
16	9-01	GC-X01								0/50ml
17	セントラル① ETRF入口	DAB-NX ①		0/50ml						
18	セントラル① ETRF出口	DAB-NX ①		0/50ml						
19	セントラル② ETRF入口	DAB-NX ②		0/50ml						
20	セントラル② ETRF出口	DAB-NX ②		0/50ml						
21	セントラル③ ETRF入口	DAB-NX ③						0/50ml		
22	セントラル③ ETRF出口	DAB-NX ③						0/50ml		
23	セントラル④ ETRF入口	DAB-NX ④						0/50ml		
24	セントラル④ ETRF出口	DAB-NX ④						0/50ml		
25	東3-03号機	GC-X01	0/50ml							
26	東3-04号機	GC-X01	0/50ml							
27	東3-05号機	GC-X01	0/50ml							
28	東3-06号機	GC-X01	0/50ml							
29	東3-07号機	GC-X01				0/50ml				
30	東3-08号機	DCS-100NX●				0/50ml				
31	東3-09号機	DCS-100NX●				0/50ml				
32	東3-10号機	DCS-100NX				0/50ml				
33	東4-09号機	DCS-100NX			0/50ml					
34	東セントラル② ETRF入口	東DAB-NX ②			0/50ml					
35	東セントラル② ETRF出口3	東DAB-NX ②			0/50ml					
36	東セントラル② ETRF出口4	東DAB-NX ②			0/50ml					