

2021年08月ET測定結果

採取日		測定部位	2021/8/3	2021/8/4	2021/8/6	2021/8/9	2021/8/16	2021/8/18	2021/8/20	2021/8/23	2021/8/27	2021/8/30
測定場所		曜日	火	水	金	月	月	水	金	月	金	月
EU/mL		担当(本館)		尾ノ上	稲富	森川	石井	山田	尾ノ上	山田	稲富	稲富
		担当(東館)	都	武田								
1	2-05	GC-X01		<0.001								
2	2-06	GC-X01		<0.001								
3	2-07	GC-X01		<0.001								
4	2-08	GC-X01		<0.001								
5	2-09	GC-X01			<0.001							
6	2-10	GC-X01			<0.001							
7	2-11	GC-X01			<0.001							
8	2-12	GC-X01			<0.001							
9	2-13	GC-X01						<0.001				
10	2-14	GC-X01						<0.001				
11	3-13	DBB-100NX						<0.001				
12	3-14	DBB-100NX						<0.001				
13	8-01	DCS-100NX										<0.001
14	8-02	DCS-100NX										<0.001
15	8-03	DCS-100NX										<0.001
16	8-04	DCS-100NX										<0.001
17	病棟①	DBB-100NX							<0.001			
18	病棟②	DBB-100NX							<0.001			
19	セントラル① ETRF出口	DAB-NX ①								<0.001		
20	セントラル② ETRF入口	DAB-NX ②								<0.001		
21	RO① モジュール	MIZ ①				<0.001	生菌測定					
22	RO② モジュール	MIZ ②				<0.001	生菌測定					
23	RO① タンク後	MIZ ①				<0.001						
24	RO② タンク後	MIZ ②				<0.001						
25	RO① EUF(奥側)	MIZ ①									<0.001	
26	RO① EUF(手前側)	MIZ ①									<0.001	
27	RO② EUF(奥側)	MIZ ②									<0.001	
28	RO② EUF(手前側)	MIZ ②									<0.001	
29	東1-09号機	GC-X01		<0.001								
30	東1-10号機	GC-X01		<0.001								
31	東2-01号機	DBB-27		<0.001								
32	東2-02号機	DCS-100NX		<0.001								
33	東4-07号機	DCS-100NX	<0.001									
34	東セントラル② ETRF入口	東DAB-NX ②	<0.001									
35	東セントラル② ETRF出口3	東DAB-NX ②	<0.001									
36	東セントラル② ETRF出口4	東DAB-NX ②	<0.001									
合計		36	4	8	4	4	0	4	2	2	4	4

2021年08月生菌測定結果

採取日		装置	2021/8/3	2021/8/4	2021/8/6	2021/8/9	2021/8/16	2021/8/18	2021/8/20	2021/8/23	2021/8/27	2021/8/30
計測日		2021/9/7	2021/8/10	2021/8/11	2021/8/13	2021/8/16	2021/8/23	2021/8/25	2021/8/27	2021/8/30	2021/9/3	2021/9/6
培養温度		25～30℃	火	水	金	月	月	水	金	月	金	月
測定場所		担当（本館）		尾ノ上	稲富	森川	石井	山田	尾ノ上	山田	稲富	稲富
		担当（東棟）	都	武田								
1	2-05	GC-X01		0/50ml								
2	2-06	GC-X01		0/50ml								
3	2-07	GC-X01		0/50ml								
4	2-08	GC-X01		0/50ml								
5	2-09	GC-X01			0/50ml							
6	2-10	GC-X01			0/50ml							
7	2-11	GC-X01			0/50ml							
8	2-12	GC-X01			0/50ml							
9	2-13	GC-X01						0/50ml				
10	2-14	GC-X01						0/50ml				
11	3-13	DBB-100NX						0/50ml				
12	3-14	DBB-100NX						0/50ml				
13	8-01	DCS-100NX										0/50ml
14	8-02	DCS-100NX										0/50ml
15	8-03	DCS-100NX										0/50ml
16	8-04	DCS-100NX										0/50ml
17	病棟①	DBB-100NX							0/50ml			
18	病棟②	DBB-100NX							0/50ml			
19	セントラル① ETRF出口	DAB-NX ①								0/50ml		
20	セントラル② ETRF入口	DAB-NX ②								0/50ml		
21	RO① モジュール	MIZ ①				1/50ml	8/50ml					
22	RO② モジュール	MIZ ②				4/50ml	1/50ml					
23	RO① タンク後	MIZ ①				0/50ml						
24	RO② タンク後	MIZ ②				0/50ml						
25	RO① EUF(奥側)	MIZ ①									0/50ml	
26	RO① EUF(手前側)	MIZ ①									0/50ml	
27	RO② EUF(奥側)	MIZ ②									0/50ml	
28	RO② EUF(手前側)	MIZ ②									0/50ml	
29	東1-09号機	GC-X01		0/50ml								
30	東1-10号機	GC-X01		0/50ml								
31	東2-01号機	DBB-27		0/50ml								
32	東2-02号機	DCS-100NX		0/50ml								
33	東4-07号機	DCS-100NX	0/50ml									
34	東セントラル② ETRF入口	東DAB-NX ②	0/50ml									
35	東セントラル② ETRF出口3	東DAB-NX ②	0/50ml									
36	東セントラル② ETRF出口4	東DAB-NX ②	0/50ml									
合計		38	4	8	4	4	2	4	2	2	4	4