

2020年12月生菌測定結果

採取日	装置	2019/12/30	2019/12/31	2020/12/1	2020/12/5	2020/12/7	2020/12/9	2020/12/10	2020/12/21	2020/12/23	2020/12/25	2020/12/28
計測日	2021/1/4	2020/1/6	2020/1/7	2020/12/8	2020/12/12	2020/12/14	2020/12/16	2020/12/17	2020/12/28	2020/12/30	2021/1/1	2021/1/4
培養温度	25~30℃	月	火	火	土	月	水	木	月	水	金	月
測定場所	担当(本館)	豊田	渡辺			吉岩	光山	中山	山田	森川	中山	大塚
	担当(東棟)			都	武田	都						
1	7-01	DCS-100NX					0/50ml					
2	7-02	DCS-100NX					0/50ml					
3	7-03	DCS-100NX					0/50ml					
4	7-04	DCS-100NX					0/50ml					
5	7-05	DCS-100NX						0/50ml				
6	7-06	DCS-100NX						0/50ml				
7	7-07	DCS-100NX						0/50ml				
8	7-08	DCS-100NX						0/50ml				
9	7-09	DCS-100NX							0/50ml			
10	7-10	DCS-100NX							0/50ml			
11	7-11	DCS-100NX							0/50ml			
12	7-12	DCS-100NX							0/50ml			
13	7-13	DCS-100NX								0/50ml		
14	7-14	DCS-100NX								0/50ml		
15	7-15	DCS-100NX								0/50ml		
16	7-16	DCS-100NX								0/50ml		
17	9-09	GC-X01		0/50ml								
18	9-10	GC-X01		0/50ml								
19	9-11	GC-X01		0/50ml								
20	9-12	GC-X01		0/50ml								
21	セントラル③ ETRF入口	DAB-NX ③				0/50ml						
22	セントラル③ ETRF出口	DAB-NX ③				0/50ml						
23	セントラル④ ETRF入口	DAB-NX ④				0/50ml						
24	セントラル④ ETRF出口	DAB-NX ④				0/50ml						
25	RO① モジュール	MIZ ①							10/50ml			
26	RO② モジュール	MIZ ①							15/50ml			
27	RO① タンク後	MIZ ②							0/50ml			
28	RO② タンク後	MIZ ②							0/50ml			
29	RO① EUF(奥側)	MIZ ①	0/50ml									0/50ml
30	RO① EUF(手前側)	MIZ ①	0/50ml									0/50ml
31	RO② EUF(奥側)	MIZ ②	0/50ml									0/50ml
32	RO② EUF(手前側)	MIZ ②	0/50ml									0/50ml
33	東3-11号機	DCS-100NX			0/50ml							
34	東3-12号機	DBB-27			0/50ml							
35	東4-10号機	DCS-100NX			0/50ml							
36	東予備機	DCS-100NX			0/50ml							
37	東RO モジュール	東MIZ		28/50ml		0/50ml						
38	東RO タンク後	東MIZ		0/50ml								
39	東RO EUF①	東MIZ		0/50ml								
40	東RO EUF②	東MIZ		0/50ml								
	合計	45	4	4	4	4	5	4	4	4	4	4